



Republika e Kosovës
Republika Kosova-Republic of Kosovo
Qeveria - Vlada - Government

Ministria e Shëndetësisë – Ministarstvo Zdravstva – Ministry of Health

Zyra e Sekretarit të Përhershëm /Ured Stalnog Sekretara /Office of the Permanent Secretary
Njësia për koordinimin e Projektit të Shëndetësisë Kosovare/ Jedinica za koordinaciju Projekta Zdravstva za Kosovo/ Coordination Unit for Kosovo Health Project

CONTRACT AWARD NOTICE

Kosovo Health Project (KHP)

Project ID: P147402

Reference No. A.2.1

CONTRACTING AUTHORITY: Ministry of Health

OBJECT OF THE CONTRACT: Procurement of Medical Equipment for MCH for hospitals

TYPE OF PROCEDURE: International Competitive Bidding (ICB)

CONTRACT AWARD CRITERIAS: Lowest price for responsible bidder

NAME OF BIDDER'S THAT SUBMITTED A BID and BIDD PRICE'S AS READ OUT:

Bidder	City	Lot Description	Price (Euro)
1 Consortium KOMED and INTREX d.o.o.	Prishtine, Novi Sad	Lot 1. Life support & monitoring equipment	974,510.00
		Lot 2. Diagnostic equipment	244,580.00
		Lot 3. Laboratory & sterilization equipment	-
		Lot 4. Surgical instruments & small med. equip.	263,269.00
		Lot 5. Ophthalmology equipment	27,470.00
2 YESSPHARMA	Prishtine	Lot 1. Life support & monitoring equipment	839,882.00
		Lot 2. Diagnostic equipment	310,968.00
		Lot 3. Laboratory & sterilization equipment	-
		Lot 4. Surgical instruments & small med. equip.	-
		Lot 5. Ophthalmology equipment	-
3 MEDBRAND	Prishtine	Lot 1. Life support & monitoring equipment	-
		Lot 2. Diagnostic equipment	297,900.00
		Lot 3. Laboratory & sterilization equipment	-
		Lot 4. Surgical instruments & small med. equip.	-
		Lot 5. Ophthalmology equipment	-
4. TBS	Trieste	Lot 1. Life support & monitoring equipment	569,686.00
		Lot 2. Diagnostic equipment	277,594.00
		Lot 3. Laboratory & sterilization equipment	166,168.50
		Lot 4. Surgical instruments & small med. equip.	-
		Lot 5. Ophthalmology equipment	-

5.	AGANI	Prishtine	Lot 1. Life support & monitoring equipment	-
			Lot 2. Diagnostic equipment	158,627.00
			Lot 3. Laboratory & sterilization equipment	-
			Lot 4. Surgical instruments & small med. equip.	-
			Lot 5. Ophthalmology equipment	-
6.	KEIS PHARMACEUTICAL	Prishtine	Lot 1. Life support & monitoring equipment	-
			Lot 2. Diagnostic equipment	289,940.00
			Lot 3. Laboratory & sterilization equipment	99,670.00
			Lot 4. Surgical instruments & small med. equip.	-
			Lot 5. Ophthalmology equipment	79,370.00
7.	MADE-KOS	Prishtine	Lot 1. Life support & monitoring equipment	-
			Lot 2. Diagnostic equipment	328,000.00
			Lot 3. Laboratory & sterilization equipment	-
			Lot 4. Surgical instruments & small med. equip.	-
			Lot 5. Ophthalmology equipment	-
8.	Consortium OPHARM and MEDICA	Prishtine	Lot 1. Life support & monitoring equipment	777,777.00
			Lot 2. Diagnostic equipment	-
			Lot 3. Laboratory & sterilization equipment	-
			Lot 4. Surgical instruments & small med. equip.	-
			Lot 5. Ophthalmology equipment	-
9.	VITECH	Prishtine	Lot 1. Life support & monitoring equipment	740,985.00
			Lot 2. Diagnostic equipment	194,030.00
			Lot 3. Laboratory & sterilization equipment	-
			Lot 4. Surgical instruments & small med. equip.	272,160.00
			Lot 5. Ophthalmology equipment	24,300.00
10.	KOSOVAMED	Prishtine	Lot 1. Life support & monitoring equipment	-
			Lot 2. Diagnostic equipment	498,750.00
			Lot 3. Laboratory & sterilization equipment	-
			Lot 4. Surgical instruments & small med. equip.	-
			Lot 5. Ophthalmology equipment	-

REJECTED BIDDER'S:

1. Consortium KOMED and INTREX d.o.o.

The Consortium "KOMED" and "INTREX" is bidding for four (4) lots: Lot 1, Lot 2, Lot 4 and Lot 5

The Evaluation Committee concluded that bidder has major deviation in regard to the technical specification for specific items for all lots, therefore is not technical responsible bid.

Lot 1 - Life support and monitoring equipment

LOT 1/.Item 2.2. There is no catalog or description to compare mandatory technical specification;

LOT 1/.Item 3. Mandatory T.S. required: *Tabletop 12-channel electrocardiographic monitor (ECG) for adults, children and neonates.* The firm has offered 6-channel ECG

LOT 1/.Item 3. Mandatory T.S. required: *Tabletop: LCD or TFT colour display, diagonal ≥ 7 ".*

The bidder/consortium has offered 5.6" (inch).

LOT 1/.Item 3. The bidder/consortium has not presented specification in regard to the Additional consumables and accessories for each unit.

LOT1/Item 5. Mandatory T.S. required integrated printer. The bidder/consortium has no specification in regard to the integrated printer.

LOT1/Item 6 and LOT1/Item 7. Mandatory T.S. required: *Equipped with rechargeable batteries*. The bidder/consortium has not presented equipment with rechargeable batteries;

LOT1/Item 10.1. Mandatory T.S. required: *Stainless steel stand and structure, a least 20 kg capacity*. The bidder/consortium has offered Stainless steel stand and structure till 15 kg.

LOT1/Item 10.2. There are no specification at all offered by the bidder/consortium.

LOT1/Item 11.2. Mandatory T.S. required: *Controls and ranges, minimum set :(tidal volume 5-2000 ml*. The bidder/consortium offered (0-999ml)

LOT1/Item 12. There are no specification at all offered by the bidder/consortium.

Lot 2 Diagnostic equipment

LOT 2/Item 1.1. There is no catalog or description to compare mandatory technical specification;

LOT 2/Item 1.2. There is no specific probe offered as well no specification to the printers as required under part of “Additional consumables and accessories for each unit”.

LOT 2/Item 1.3. There is no specific probe offered as well no specification to the printers as required under part of “Additional consumables and accessories for each unit”.

Lot 4 Surgical Instruments and small equipment

LOT 4. Five items (4, 5, 6.1, 6.2, and 15) were not offered;

LOT 4. For some of items were not offered catalogs, or the type of offered devices was not corresponding to the T.S.

Lot 5 Ophthalmology equipment

LOT 5/ Item 1 and Item 2. Operational characteristics as well electrical characteristics are not corresponding to the T.S.

2. YESSPHARMA

The bidder is bidding for two (2) lots: Lot 1 and Lot 2.

The Evaluation Committee concluded that bidder has major deviation in regard to the technical specification for specific items under Lot 2 Diagnostic equipment, therefore is not technical responsible bid.

Lot 2 Diagnostic equipment

LOT 2/Item 1.1, Item 1.2 and Item 1.3.

Mandatory T.S. required: *Equipped with rechargeable batteries (integrated or by external UPS module):*
a) *Working autonomy ≥ 1 h*. The firm has offered 30 min.

Mandatory T.S. required: *Hard disk ≥ 1 TB*. The firm has offered 550 GB

LOT 2/Item 1.1.

Mandatory T.S. required: *Equipped with rechargeable batteries (integrated or by external UPS module):*

b) *Visual and audible alarm in case of low battery*. There is no specification.

c) *Automatic switch between line and battery powered modes*. There is no specification.

d) *Integrated battery charger*. There is no specification.

Protections against over-voltage and over-current line conditions. There is no specification.

LOT 2/Item 1.1.

Mandatory T.S. required: *Dynamic range (gray scale compression) ≥ 160 dB*

The firm has offered (36-96dB) with 3 OR 6 STEPS)

LOT 2/Item 1.1 and Item 1.3

Mandatory T.S. required: *The following probes must be included in the offer: CONVEX array, ob-gyn/abdominal/adult-pediatric/vascular at least triple frequency, bandwidth approximately 3-5.5 Mhz, including the 3.5 Mhz frequency.*

The firm has offered 2-5 Mhz

LOT 2/Item 1.2

Mandatory T.S. required: *The equipment shall be compatible with at least all types of standard ob/gyn, neonatal and general use transducers (linear, convex, microconvex, endocavity, trans-vaginal, phased):*

a) *PHASED array, pediatric-neonatal/cardio/cephalic/trans-craneal, bandwidth approximately 5-10 MHz.*

The firm has offered 3-7 Mhz

3. MEDBRAND

The bidder is bidding for Lot 2.

The Evaluation Committee concluded that bidder has major deviation in regard to the technical specification for specific items under Lot 2 Diagnostic equipment, therefore is not technical responsible bid.

Lot 2 Diagnostic equipment

LOT 2/Item 1.1.

Mandatory T.S. required: *Equipped with rechargeable batteries (integrated or by external UPS module):*

b) *Visual and audible alarm in case of low battery. There is no specification.*

c) *Automatic switch between line and battery powered modes. There is no specification.*

d) *Integrated battery charger. There is no specification.*

Protections against over-voltage and over-current line conditions. There is no specification.

LOT 2/Item 1.1.

Mandatory T.S. required: *Hard disk ≥ 1 TB.*

The firm has offered 550 GB

LOT 2/Item 1.1.

Mandatory T.S. required: *The following probes must be included in the offer: CONVEX array, ob-gyn/abdominal/adult-pediatric/vascular at least triple frequency, bandwidth approximately 3-5.5 Mhz, including the 3.5 Mhz frequency.*

The firm has offered 1.4-5 Mhz

LOT 2/Item 1.1.

Mandatory T.S. required: *The equipment shall be compatible with at least all types of standard ob/gyn, neonatal and general use transducers (linear, convex, microconvex, endocavity, trans-vaginal, phased):*

a) *PHASED array, pediatric-neonatal/cardio/cephalic/trans-craneal, bandwidth approximately 5-10 MHz.*

The firm has offered 3.1-8.8 Mhz

4. TBS

The bidder is bidding for three (3) lots: Lot 1, Lot 2 and Lot 3;

The Evaluation Committee concluded that bidder has major deviation in regard to the technical specification for specific items under LOT 1 Life support and monitoring equipment, Lot 2 Diagnostic equipment and Lot 3 Laboratory and sterilization equipment, therefore is not technical responsible bid.

LOT 1 Life support and monitoring equipment

Lot 1/Item 1.

Mandatory T.S. required: *Adult, pediatric and neonatal modes.* The firm/bidder has no specification.

Lot 1/Item 1.

Mandatory T.S. required: *Integrated or external printer*. The firm/bidder has no specification.

Lot 1/Item 3.

Mandatory T.S. required: *LCD or TFT colour display, diagonal $\geq 7''$, with visualization of analogical curves and alphanumeric values*. The firm/bidder has offered 4.3" inch.

Lot 1/Item 3.

Mandatory T.S. required: *Maximum discharge energy setting with mono-phase wave not less than 320 J*. The firm/bidder has offered 200 J

Lot 1/Item 4.1, Item 6 and Item 7.

The offered equipment are not compliance with Electrical characteristics

Lot 2 Diagnostic equipment

Lot 2/Item 2.

Mandatory T.S. required: *Monitor (LED, LCD or TFT flat color monitor): b) diagonal ≥ 17 inches*. The firm/bidder has offered 15 inches;

Mandatory T.S. required: *Hard disk ≥ 1 TB*. The firm/bidder has offered Hard disk 320 GB;

Mandatory T.S. required: *the following probes must be included in the offer and shall be compatible to applicable 3D/4D modes: a) PHASED array, pediatric-neonatal/cardio/cephalic/trans-craneal, bandwidth approximately 5-10 MHz*. The firm/bidder has offered 2-8 MHz

Lot 2/ Item 1.

Mandatory T.S. required: *the following Monitor must be included in the offer (LED, LCD or TFT flat color monitor): b) diagonal ≥ 17 inches*. The firm/bidder has offered 15 inches;

Mandatory T.S. required: the following Frequency range 1-15 MHz or wider must be included in the offer.

Lot 3 Laboratory and sterilization equipment

Lot 3/Item 1.

Mandatory T.S. required:

- *Temperature sensor and external display, showing set and working temperatures.*
- *Self-check and malfunction alarms.*

The firm/bidder has no specification.

Lot 3/Item 1.2

Mandatory T.S. required:

- *Internal chamber equipped with multiple rails (at least 4) for shelve positioning at various heights*. The firm/bidder has offered standard 2 shelves.
- *Internal capacity approximately 55 l*. The firm/bidder has offered 42 L

Lot 3/Item 1.3

Mandatory T.S. required: *(between 450 and 550 nm or other certified method) or multi-wavelength spectral reflectance technology*. The firm/bidder has offered 450 and 475nm.

5. AGANI

The bidder is bidding for Lot 2;

The Evaluation Committee concluded that bidder has major deviation in regard to the technical specification for specific items under Lot 2 Diagnostic equipment, therefore is not technical responsible bid.

Lot 2 Diagnostic equipment

LOT 2/Item 1.1.

Mandatory T.S. required: *Equipped with rechargeable batteries (integrated or by external UPS module):*

a) Working autonomy ≥ 1 h. There is no specification.

- b) *Visual and audible alarm in case of low battery.* There is no specification.
- c) *Automatic switch between line and battery powered modes.* There is no specification.
- d) *Integrated battery charger.* There is no specification.

LOT 2/Item 1.1. ; Item 1.2 and Item 1.3.

Mandatory T.S. required: *Hard disk ≥ 1 TB.* The firm/bidder has offered 320 GB;

LOT 2/Item 1.1. ; Item 1.2 and Item 1.3.

Mandatory T.S. under Additional consumables and accessories required: *1 color thermal printers for medical ultrasound applications.* There is no specification.

6. KEIS PHARMACEUTICAL

The bidder is bidding for three (3) lots: Lot 2, Lot 3 and Lot 5;

The Evaluation Committee concluded that bidder has major deviation in regard to the technical specification for specific items under Lot 2 Diagnostic equipment, therefore is not technical responsible bid.

Lot 2 Diagnostic equipment

LOT 2/Item 1.1. ; Item 1.2 and Item 1.3.

Mandatory T.S. required: *Equipped with rechargeable batteries (integrated or by external UPS module): Working autonomy ≥ 1 h.* There is no specification.

LOT 2/Item 1.1. ; Item 1.2 and Item 1.3.

Mandatory T.S. under Additional consumables and accessories required: *1 one B/W and 1 one color thermal printers for medical ultrasound applications.* There is no specification.

LOT 2/Item 1.1 and Item 1.3.

The Mandatory T.S. required: *different probes must be included in the offer: CONVEX, LINEAR etc.* There is no specification.

7. VITECH

The bidder is bidding for four (4) lots: Lot 1, Lot 2, Lot 4 and Lot 5;

The Evaluation Committee concluded that bidder has major deviation in regard to the technical specification for specific items for all lots, therefore is not technical responsible bid.

LOT 1 Life support and monitoring equipment

LOT /Item 1.

Mandatory T.S. required: *LCD or TFT color; diagonal ≥ 10 inches.* The firm/bidder has offered 7 inches;

LOT /Item 1.

Mandatory T.S. required: *Up to 12 standard leads ECG module.* The firm/bidder has offered only 3 leads;

LOT /Item 1.

Mandatory T.S. required: *Bandwidth at least 0,05-250 Hz.* The firm/bidder has offered 0.5Hz-40Hz.

LOT /Item 4.

Mandatory T.S. required: *Ultrasound transducer working frequency inside the range (approximate) from 1.0 to 8 MHz.* The firm/bidder has offered 2 MHz.

Lot 2 Diagnostic equipment

LOT 2/Item 1.1. ; Item 1.2 and Item 1.3.

Mandatory T.S. required: *the following Monitor must be included in the offer (LED, LCD or TFT flat color monitor): b) diagonal ≥ 17 inches.* The firm/bidder has offered 15 inches;

LOT 2/Item 1.1. ; Item 1.2 and Item 1.3.

Mandatory T.S. required: the following *Hard disk ≥ 1 TB* must be included in the offer. The firm/bidder has offered Hard disk 320 GB;

LOT 2/Item 1.1

Mandatory T.S. required: *to be equipped with rechargeable batteries (integrated or by external UPS module):*

- a) *working autonomy ≥ 1 h;*
- b) *visual and audible alarm in case of low battery;*
- c) *automatic switch between line and battery powered modes;*

The firm/bidder has no specification for all of them.

Lot 4 Surgical Instruments and small equipment

LOT 4/ items (4, 5, 6.1, 6.2) were not offered;

LOT 4/ to the many items were not offered catalogs, or the type of offered devices have several deviations and not corresponding to the T.S. (i.e. 3, 7.1, 8, 10, 12.2, 13, 16, 17.1, 17.2, 17.3, 19)

Lot 5 Ophthalmology equipment

LOT 5/ Item 1 and Item 2.

Operational characteristics as well electrical characteristics are not corresponding to the T.S.

8. KOSOVAMED

The bidder is bidding for Lot 2;

The Evaluation Committee concluded that bidder has major deviation in regard to the technical specification for specific items under Lot 2 Diagnostic equipment, therefore is not technical responsible bid.

Lot 2 Diagnostic equipment

LOT 2/Item 1.1, Item 1.2 and Item 1.3.

Mandatory T.S. required: *Equipped with rechargeable batteries (integrated or by external UPS module):*

Working autonomy ≥ 1 h. The firm has offered 40 min.

Mandatory T.S. required: *Hard disk ≥ 1 TB.* The firm/bidder has offered 512 GB;

LOT 2/Item 1.1, Item 1.2 and Item 1.3.

Mandatory T.S. under Additional consumables and accessories required: *1 (one) B/W and 1 (one) color thermal printers for medical ultrasound applications;* There is no specification.

CONTRACT(S) AWARD:

Lot 1- Life support and monitoring equipment

Joint Venture: N.T.P. Opharm, Prishtine and N.T.P. Medica, Prishtine

The contract price: 777,777.00 Euro

Lot 2 Diagnostic equipment

N.SH.T. MADE-KOS, Prishtine

The contract price: 328,000.00 Euro

Lot 3 - Laboratory and sterilization equipment

KEIS (Kosova Export Import Supply Pharmaceuticals) Sh.p.k. Prishtine

The contract price: 99,360.00 Euro

Lot 5 - Ophthalmology equipment

KEIS (Kosova Export Import Supply Pharmaceuticals) Sh.p.k. Prishtine

The contract price: 79,370.00 Euro

Lot 4 - Surgical instruments and small medical equipment

This lot is cancelled while two bidders are not substantially responsive in regard to the technical specifications requirements;